



Certificate of Accreditation

International Accreditation Japan (IAJapan) hereby accredits the following conformity assessment body as a testing laboratory of ASNITE accreditation program.

Accreditation Identification: ASNITE 0032 Testing

Name of Conformity Assessment Body: Type Test Group Verification Management Division,
Japan Electric Meters Inspection Corporation

Name of Legal Entity: Japan Electric Meters Inspection Corporation

Location of Conformity Assessment Body: 15-7, 4-chome, Shibaura, Minato-ku, Tokyo 108-0023,
JAPAN

Scope of Accreditation: as the following pages

Accreditation Requirement: ISO/IEC 17025:2017*

* The relevant accreditation requirements described in the Accreditation Scheme Document for ASNITE-T (G) are also applied.

Effective Date of Accreditation: 2023-09-19

Expiry Date of Accreditation: 2027-09-18

Date of Initial Accreditation: 2009-07-23

A handwritten signature in black ink, appearing to read 'K. Saito', is written over a horizontal line.

SAITO Kazunori

Chief Executive, International Accreditation Japan (IAJapan)

National Institute of Technology and Evaluation

-
- International Accreditation Japan (IAJapan) is a laboratory accreditation body which has signed MRAs of ILAC (International Laboratory Accreditation Cooperation) and APAC (Asia Pacific Accreditation Cooperation).
 - MRA requirements are, in addition to relevant international standards and guides, requirements for participation in proficiency testing programs, surveillance and reassessment, and the policy for the traceability of measurement for MRA purpose.
 - This laboratory fulfills ISO/IEC 17025:2017 General requirements for the competence of testing and calibration laboratories. This accreditation means this laboratory meets both the technical competence requirements and management system requirements that are necessary for it to consistently deliver technically valid test results and calibrations (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).
 - The latest accreditation information is publicly available on IAJapan Website as an accreditation certificate.

Name of Laboratory : Type Test Group Verification Management Division,
 Japan Electric Meters Inspection Corporation
 Address : 15-7, 4-chome, Shibaura, Minato-ku, Tokyo 108-0023, Japan
 Conformity : Working within Accredited Scope of Type Test Group
 Assessment Activities Verification Management Division

< Verification Management Division Type Test Group's Scope of Accreditation >

Effective Date of Accreditation: 2023-09-19				
Materials or Products Tested	Test Type (Testing Method (s))	Component, Parameter or Characteristic Tested	Testing Method Standard(s)	Notices
Normal static meters for active energy and Precision static meters for active energy	Test of electrical performance etc. specified in Article 681 or Article 725 of Regulation for Verification and Inspection of Specified Measuring Instruments based on the Measurement Act	Electrical performance	Following test methods specified in JIS C1271-2 referred to in Article 725 of Regulation for Verification and Inspection of Specified Measuring Instruments based on the Measurement Act JIS C1271-2 7.2.1, 7.2.2, 7.2.3, 7.2.4, 7.2.5, 7.2.6, 7.3.1, 7.3.2, 7.3.3, 7.3.4, 7.3.5, 7.3.7, 7.3.8, 7.3.9, 7.3.10, 7.3.11, 7.3.12, 7.3.13, 7.3.14.1, 7.3.14.2, 7.3.15, 7.3.16, 7.4.2, 7.4.3, 7.4.4, 7.4.5, 7.4.6, 7.4.7, 7.4.9 and 7.4.10	-
	Test of mechanical performance etc. specified in Article 681 or Article 725 of Regulation for Verification and Inspection of Specified Measuring Instruments based on the Measurement Act	Mechanical performance	Following test methods specified in JIS C1271-2 referred to in Article 725 of Regulation for Verification and Inspection of Specified Measuring Instruments based on the Measurement Act JIS C1271-2 7.3.6, 7.4.11.1, 7.4.11.2, 7.4.12, 7.4.14, 7.9, 7.10, 7.11, 7.12 and 7.13	
	Test of temperature rise test of the load current conductor and terminal specified in Article 681 or Article 725 of Regulation for Verification and Inspection of Specified Measuring Instruments based on the Measurement Act	Temperature rise test of the load current conductor and terminal	Following test methods specified in JIS C1271-2 referred to in Article 725 of Regulation for Verification and Inspection of Specified Measuring Instruments based on the Measurement Act JIS C1271-2 7.8	

Test of insulation performance specified in Article 681 or Article 725 of Regulation for Verification and Inspection of Specified Measuring Instruments based on the Measurement Act	Insulation performance	Following test methods specified in JIS C1271-2 referred to in Article 725 of Regulation for Verification and Inspection of Specified Measuring Instruments based on the Measurement Act JIS C1271-2 7.4.8, 7.5.1 and 7.5.2
Test of weather resistance etc. specified in Article 681 or Article 725 of Regulation for Verification and Inspection of Specified Measuring Instruments based on the Measurement Act	Weather resistance	Following test methods specified in JIS C1271-2 referred to in Article 725 of Regulation for Verification and Inspection of Specified Measuring Instruments based on the Measurement Act JIS C1271-2 7.4.13.1, 7.4.13.2, 7.4.13.3, 7.4.13.4, 7.4.13.5, 7.4.13.6, 7.4.13.7, 7.4.13.8, 7.4.13.9, 7.4.13.10 and 7.7
Test of material specified in Article 681 or Article 725 of Regulation for Verification and Inspection of Specified Measuring Instruments based on the Measurement Act	Material	Following test methods specified in JIS C1271-2 referred to in Article 725 of Regulation for Verification and Inspection of Specified Measuring Instruments based on the Measurement Act JIS C1271-2 7.6 a) and 7.6 b)

(End of Certificate)