



Certificate of Accreditation

International Accreditation Japan (IAJapan) hereby accredits the following conformity assessment body as a testing laboratory of ASNITE accreditation program.

Accreditation Identification: ASNITE 0056 Testing

Name of Conformity Assessment Body: Yaizu Laboratory, General Incorporated Foundation
Shizuoka Chemical Analysis Center

Name of Legal Entity: General Incorporated Foundation Shizuoka Chemical
Analysis Center

Location of Conformity Assessment Body: 1-1, Shiotsu, Yaizu-shi, Shizuoka 425-0085, JAPAN

Scope of Accreditation: As the following pages

Accreditation Requirement: ISO/IEC 17025:2017*

* The relevant accreditation requirements described in the Accreditation Scheme Document for ASNITE-T (E) are also applied.

Effective Date of Accreditation: 2023-09-30

Expiry Date of Accreditation: 2027-09-29

Date of Initial Accreditation: 2011-06-29

ISHIGE Hiromi

ISHIGE Hiromi

Chief Executive, International Accreditation Japan (IAJapan)

National Institute of Technology and Evaluation

-
- International Accreditation Japan (IAJapan) is a laboratory accreditation body which has signed MRAs of ILAC (International Laboratory Accreditation Cooperation) and APAC (Asia Pacific Accreditation Cooperation).
 - MRA requirements are, in addition to relevant international standards and guides, requirements for participation in proficiency testing programs, surveillance and reassessment, and the policy for the traceability of measurement for MRA purpose.
 - This laboratory fulfills ISO/IEC 17025:2017 General requirements for the competence of testing and calibration laboratories. This accreditation means this laboratory meets both the technical competence requirements and management system requirements that are necessary for it to consistently deliver technically valid test results and calibrations (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).
 - The latest accreditation information is publicly available on IAJapan Website as an accreditation certificate.

(Attachment)

Name of Laboratory: Yaizu Laboratory, General Incorporated Foundation Shizuoka Chemical Analysis Center
 Address of Laboratory: 1-1, Shiotsu, Yaizu-shi, Shizuoka 425-0085, JAPAN
 Work to Carry Out: Control of management system, Service to the customer, Review of requests, Sampling, Sample storage, Test execution, Ensuring the validity of results, Reporting of results

Accreditation Scope			Testing Items	Test Methods	Effective Date of Accreditation
Category	Sub-Category	Measurement Techniques			
Chemical Products	Water	Ion electrode method (Including pH measurement)	pH / Water Source, Raw Water, Treated Water (including sampling)	Standard Methods for the Examination of Water, II-3 9.2 (JWWA:2020)	2023-09-30
			pH / Raw Water, Treated Water, Tap Water, Drinking Water (including sampling)	MHLW Notification No.261:2003 (Revised to MOE Notification No.25:2025) Appended Table 31	2023-09-30

[NOTE]

MOE: Ministry of the Environment

JWWA: Japan Water Works Association

(End of Attachment)